

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993
CONCERNING MEDICAL DEVICES**



MANUFACTURER: NAME: Guangdong Transtek Medical Electronics Co.,Ltd.
ADDRESS: Zone A, No.105 ,Dongli Road , Torch Development District,
Zhongshan, Guangdong, China

MEDICAL DEVICE: BLOOD PRESSURE MONITORS: TMB-1585-BT

CLASSIFICATION - ANNEX IX: CLASS IIA, RULE 10

CONFORMITY ASSESSMENT ROUTE: MDD ANNEX II

WE, THE MANUFACTURER, HEREWITH DECLARE THAT THE STATED MEDICAL DEVICES MEET THE TRANSPOSITION INTO NATIONAL LAW, THE PROVISIONS OF COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 CONCERNING MEDICAL DEVICES; INCLUDING, AT 21 MARCH 2010, THE AMENDMENTS BY COUNCIL DIRECTIVE 2007/47/EEC. ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURER.

STANDARDS APPLIED: SEE ATTACHED

NOTIFIED BODY: TÜV SÜD PRODUCT SERVICE GMBH
RIDLERSTR 65, D-80339 M NCHEN, GERMANY

IDENTIFICATION NUMBER **CE 0123**

(EC) CERTIFICATE(S): NO.G1 16 11 82800 026




EUROPEAN REPRESENTATIVE: MDSS-MEDICAL DEVICE SAFETY SERVICE GMBH
SCHIFFGRABEN ,41,30175, HANNOVER, GEMANY

START OF CE-MARKING: 2015-11-3

PLACE, DATE OF DECLARATION: ZHONGSHAN, 2017-3-2

SIGNATURE:

NAME: 
POSITION: Vice President

**DECLARATION OF CONFORMITY
TO COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993
CONCERNING MEDICAL DEVICES**

Standards applied:

| | |
|--|---|
| Risk management | EN ISO 14971:2012 |
| Labeling | EN 980:2008 |
| User manual | EN 1041: 2008 |
| General requirements for safety | EN 60601-1: 2006/IEC 60601-1:2005+A1: 2012 EN 60601-1-11:2011 |
| Non-invasive sphygmomanometers General requirements | EN ISO 81060-1:2012 EN 1060-3:1997+A2:2009 IEC/EN 80601-2-30:2009 |
| Electromagnetic compatibility | EN 60601-1-2:2007 |
| Usability | EN 60601-1-6:2010 & EN 62366:2007 |
| Software life-cycle | EN 62304:2006+AC: 2008 |
| Biological evaluation | EN ISO 10993-1:2009 EN ISO 10993-5:2009 EN ISO 10993-10:2010 |
| Clinical Investigation | MEDDEV.2.7.1: 2009 EN 1060-4: 2004 |